

**Department of Forensic Science**  
**Breath Alcohol Section**  
**Instrument Location History**  
**1/1/1990 To 1/26/2015**

*mmB*

**Instrument Serial Number: 010474**

**As of 26-Jan-15**

<b>Start Date</b>	<b>End Date</b>	<b>Agency Name</b>
21-Jan-15		DFS Central Lab
05-Aug-14	21-Jan-15	Norton PD
29-Jul-14	05-Aug-14	DFS Central Lab
10-Feb-14	29-Jul-14	Greensville County Sheriff's Office
01-Apr-13	10-Feb-14	DFS Central Lab
20-Dec-12	01-Apr-13	Virginia Beach PD
09-Nov-12	20-Dec-12	DFS Central Lab
03-Mar-11	09-Nov-12	Brunswick County SO
23-Jun-09	03-Mar-11	DFS Central Lab
07-Nov-08	23-Jun-09	Greensville County Sheriff's Office
13-May-08	07-Nov-08	DFS Central Lab

INTOX EC/IR II  
Quality Assurance Worksheet

AMS  
mmB

Instrument Serial Number 010474 Worksheet Start Date 1/21/2015  
Location Norton Police Department  
Address 618 Virginia Ave., Norton, VA 24273  
DFS Technician Heather Stanton License No. 18910  
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 704 Reference Barometer (mm HG) 708  
Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.281
0.300	0.278	0.269	0.286	Sample 2	0.281
Precision		sample min	sample max	Sample 3	0.281
0		0.281	0.281		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.092
0.100	0.093	0.090	0.096	Sample 2	0.093
Precision		sample min	sample max	Sample 3	0.092
0.001		0.092	0.093		

Dry gas standard Lot No. (with tank no.) AG400202-41

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

Instrument Serial Number

010474

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Starch

Date

1/26/15


Issuing Analyst

Michelle B. Lillie

Date

1/27/15

JPD




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>STANTON, HEATHER, M</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>18910</b>	LICENSE EXPIRES <b>10/01/2016</b>	DATE TEST CONDUCTED <b>01/21/2015</b>
TEST EQUIPMENT NUMBER <b>010474</b>		

RESULTS: TIME SAMPLE TAKEN **13:51 EST**

SAMPLE'S ALCOHOL CONTENT **0.00** GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR
----------------------

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 21 Jan 2015 at 14:44

TIN  
HHS  
mmB

## Test Results

Instrument Serial Number 010474

Test # 000711 Subject Test

Test Location 1 Department of  
Test Date 21 Jan 2015

Test Location 2 Forensic Science  
Test Time 13:45  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name STANTON  
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial N

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 21 Jan 2015 End Time 13:53

Result Time 13:51

Result Date 21 Jan 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 13:45

Data Type BLK

Sample Value 0.000

Sample Time 13:46

Data Type CHK

Sample Value 0.093

Sample Time 13:47

Data Type BLK

Sample Value 0.000

Sample Time 13:48

Data Type SUBJ

Sample Value 0.000

Sample Time 13:49

Data Type BLK

Sample Value 0.000

Sample Time 13:50

Data Type SUBJ

Sample Value 0.000

Sample Time 13:51

Data Type BLK

Sample Value 0.000

Sample Time 13:52

Standard Type Dry Gas Std

Standard Value 0.093

Standard Lot Number AG400202-41

Standard Expiration Date 01/02/2016

Tank Pressure 389

Barometric Pressure 704 mmHg

Blow Sample Number 1 Blow Duration 3.35 sec

Blow Volume 1826 cc End-of-Blow Time 13:49

Blow Sample Number 2 Blow Duration 3.12 sec

Blow Volume 1759 cc End-of-Blow Time 13:51

Tamper Evident Stamp 41142324

Test Status Code 0

Test Status Success

AMS  
mmB

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010474      Test Number: 712

Test Date: 01/21/2015      Test Time: 13:54 EST

Dry Gas Target: 0.093

Lot Number: AG400202-41      Exp Date: 01/02/2016

Tank Pressure: 384 psi      Barometric Pressure: 704 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:55
CHK	0.092	13:56
BLK	0.000	13:58
CHK	0.093	13:58
BLK	0.000	14:00
CHK	0.092	14:00

Test Status: *Success*

Calibration CRC: BE0BF310

1/18  
mmB

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010474      Test Number: 713

Test Date: 01/21/2015      Test Time: 14:04 EST

Dry Gas Target: 0.278

Lot Number: AG428003-09      Exp Date: 10/07/2016

Tank Pressure: 291 psi      Barometric Pressure: 704 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:05
CHK	0.281	14:05
BLK	0.000	14:07
CHK	0.281	14:08
BLK	0.000	14:10
CHK	0.281	14:11

Test Status: *Success*

Calibration CRC: BE0BF310